



# Declaration of Conformity

## ***dataTaker***

### Manufacturer

*Name* Thermo Fisher Scientific Australia Pty. Ltd.  
*Address* 5 Caribbean Drive  
Scoresby 3179 Australia  
*ABN* 52 058 390 917  
*Supplier Code* N1105

### Conforming Products

DT50, 51	Data Logger	CANgate	CAN Bus Interface
DT500, 505, 515	Data Logger	DT800	Data Logger
DT600, 605, 615	Data Logger	DT80, 81, ,82E 85	Data Logger
PMD-01	Panel Mount Display	DT80G, 85G	Data Logger
CEM-AD	Channel Expansion Model	CEM20	Channel Expansion Model

### EMC Conformance

Thermo Fisher Scientific Australia Pty Ltd hereby declare that the above products conform with the Electromagnetic Compatibility Standard as required under the Radio Communications Act.

*Applicable Standards* AS/NZS CISPR 22: 2004 Information technology equipment - Radio disturbance characteristics  
- Limits and methods of measurement  
AS/NZS CISPR 24: 2002 Information technology equipment - Immunity characteristics  
- Limits and methods of measurement

### Testing Organisation

*Name* RFI Industries Pty Ltd. - EMI/EMC Laboratories  
*Address* 52 - 54 Holloway Drive, Bayswater, VIC 3153 Australia

### Associated Equipment

Mains adaptors used to power the above products must conform with the provisions of EN60950, EN60742 or EN61010.

### Signatory

*Signed* ..... for and on behalf of Thermo Fisher Scientific Australia Pty. Ltd.  
Albertus Mouton - Commercial Manager